## **Analytical Method Information**

		Reporting Surrogate Duplicate		Duplicate	Matrix Spike		Blank Spike / LCS	
Analyte	MDL	Limit	%R	RPD	%R	RPD	%R	RPD
Ammonia -R8 in Water (EPA 3	50.1)							
Preservation: 02-Sulfuric Acid (H2SO4) to pH<2, Cool 4°C								
Container: 04_125mL Plastic pH <2 w/		Amount Required: 125 mL			Hold Time: 28 days			
H2SO4								
Ammonia as N	0.00240	0.0500  mg/L		20	80 - 120	20	80 - 120	